The ATLAS Tile Calorimeter performance at the LHC

Robert Calkins -NIU/NICADD
on behalf of the ATLAS Tile Calorimeter Group

Outline

- What is the Tile Calorimeter?
- How well is it operating?
- How is it calibrated?
- How precise and stable are these calibration systems?
- Can we measure luminosity with it?
- What about cosmic rays and collision events?
- How is E/p of hadrons determined, does it agree with simulation?

TileCal – A non-compensating hadronic sampling calorimeter

- Consists of 5182 steel and scintillator structured cells connected to PMTs by WLS fibers
- Approximately 10k PMTs, majority of cells are readout by two PMTs
- 0.1 x 0.1 granularity in eta-phi for first 2 layers, 0.2 x 0.1 for third layer
- Central Hadronic component of combined calorimeter, covers $|\eta|$ out to ~ 1.7

Calibration and Read-out Chain

Calibration schema in Tile Calorimeter

- Light propagates from scintillators through WLS fibers to PMTs
- Pulse from PMT is sent through a shaper that stretches it out
- This is sampled 7 times at 25ns intervals by an ADC
- Charge is reconstructed using an Optimal Filtering algorithm implemented on the hardware
- Information from one or both PMTs is used to reconstruct cell quantities such as energy and timing
- Digital Electronics have two gains
- Charge integrators read PMT current in Cs scans and during collisions

Detector Status and Monitoring

- Calibration runs taken every few days to maintain calibration, also used to spot new problems
- Data quality is monitored in ATLAS control room and offline after reconstruction has finished
- PMT voltages and temperatures are constantly monitored and are stable over long periods of time

Charge Injection System (CIS)

- Used to derive calibration constants from pulse charge (pC) to ADC
- Charge is injected into the electronics chain to test stability and linearity of the system
- Runs 10k events are taken with a fixed amplitude to study stability of response and to identify noise
- Other runs span the entire dynamic range of ADC and used to test linearity of electronics response

Cesium Source

Evolution of Cesium integrals with time

- Constants are adjusted to maintain EM scale as defined at testbeam
- System is incorporated with detector, radioactive source travels through cells
- Monthly scans are taken with and without ATLAS B-field, affect is larger (0.5%) in central barrel than in extended barrel (0.2%)
- Precision is 0.3% after cell response inter-calibration

Stability with Laser

- Runs taken much more frequently than Cs, important cross-check of PMTs response stability between Cs scans
- PMT response to laser is quite stable over several weeks – most PMTs <1%
- Gain is sensitive to HV, we can measure and correct very accurately

Timing

Time of Flight
Correction

- Timing is set using beam splash* events – all cells have significant signal
- Timing is corrected for time of flight and WLS fiber length
- TileCal has excellent timing resolution – RMS of cells = 0.5 ns

Splash events are produced by the proton beam hitting collimators located 140m away from ATLAS

Minimum Bias Integrating Current

- TileCal is equipped with integrator cards that record over $O(10^5)$ bunch crossings currents passing through PMT
- Integrator circuit stability is validated by a gain measurement system
- Integrated current is proportional to luminosity to per mil precision over an order of magnitude

EM-Scale

- A subset of modules ($\sim 11\%$) were subjected to an electron testbeam and EM scale defined with a RMS spread of 2.4%
- Rest of the modules were inter-calibrated to this subset in assembled ATLAS detector
- Uncertainties in transferring testbeam EM scale come from calibration related uncertainties:
 - 0.1% from CIS (pC $\rightarrow$ ADC) calibration
 - 0.2% from Cs gain
 - 0.5% from B-field effect during Cs runs
 - 0.3% from radial reweighing for different size scintillators
 - 1.5% from Cs updrift of 0.1% per month

Validation of EM Scale and Uniformity with Cosmics (1)

- Cosmic muon tracks are selected with momentum from 10 and 30 GeV/c
- For cells to be considered, must have >100 events passing selection – still a significant fraction of cells
- Cells in all three layers have uniform response with cosmic muons, close to expectations from MC
- Similar analysis performed with muons at testbeam

Layer	Number of cells	Fraction of cells	RMS (MeV/mm)	
			Data	MC
A	352	18%	0.060	0.049
BC	421	22%	0.046	0.043
D	316	38%	0.052	0.048

Validation of EM Scale and Uniformity with Cosmics (2)

Radial layer		A	BC	D
Cosmic muons, LB	Data	$1.28^{+0.03}_{-0.04}$	1.32 ± 0.05	1.35 ± 0.04
	MC	1.32 ± 0.04	1.35 ± 0.05	1.34 ± 0.04
	Data/MC	$0.97^{+0.01}_{-0.02}$	0.98 ± 0.02	1.01 ± 0.01
Cosmic muons, EB	Data	1.27 ± 0.06	1.29 ± 0.06	1.32 ± 0.05
	MC	1.31 ± 0.03	1.32 ± 0.06	1.34 ± 0.05
	Data/MC	0.97 ± 0.04	0.98 ± 0.03	0.99 ± 0.02
Testbeam, LB	Data	1.25 ± 0.03	1.39 ± 0.04	1.39 ± 0.03
	MC	1.30 ± 0.02	1.37 ± 0.03	1.36 ± 0.02
	Data/MC	0.96 ± 0.02	1.02 ± 0.04	1.02 ± 0.02
Double ratio $\frac{(Data/MC)_{Cosmic\ muons, LB}}{(Data/MC)_{TB, LB}}$		1.01 ± 0.03	0.96 ± 0.04	0.98 ± 0.03

- As estimator of uncertainty, double ratio is defined between $(Data/MC)_{Cosmics} / (Data/MC)_{Testbeam}$ - should = 1
- All three layers agree within errors – we've successfully propagated EM scale from testbeam to detector
- Results are 1.01, 0.96 and .08 for A,BC and D respectively
- EM scale uncertainty is quoted as largest deviation which is 4%

E/p from hadrons in Collisions

- Isolated particles showing in TileCal are selected, momentum is measured with tracking inner detector
- Excellent agreement with MC (hadron shower tuned on testbeam data)
- Tool for in-situ uniformity tests

Summary

- Tile has been working well during the first year of collisions
- Tile is operating under stable conditions
- Calibration systems are within design precisions and show excellent stability over time
- Timing has been set and validated with collisions and splash events
- Minimum bias integrator system has demonstrated its ability to follow luminosity measurements
- EM scale has been transferred from testbeam measurements and validated with cosmic muons

Backup Slides

Full Table of EM Scale Systematic Uncertainties

Systematic Uncertainties [MeV/mm] for Data and MC

Uncertainty source		Long Barrel			Extended Barrel		
		A	BC	D	A	B	D
Path	Data	± 0.016	± 0.030	± 0.019	± 0.046	± 0.030	± 0.017
	MC	± 0.006	± 0.008	± 0.013	± 0.014	± 0.015	± 0.022
	Data/MC ratio	± 0.008	± 0.016	± 0.009	± 0.033	± 0.021	± 0.019
Momentum	Data	± 0.024	± 0.034	± 0.033	± 0.037	± 0.043	± 0.044
	MC	± 0.032	± 0.042	± 0.035	± 0.020	± 0.042	± 0.044
	Data/MC ratio	± 0.008	± 0.007	± 0.004	± 0.024	± 0.005	± 0.009
Noise	Data	± 0.007	± 0.002	± 0.002	± 0.009	± 0.004	± 0.003
	MC	± 0.004	± 0.002	± 0.003	± 0.003	± 0.002	± 0.002
	Data/MC ratio	± 0.002	± 0.000	± 0.001	± 0.005	± 0.001	± 0.000
Truncation	Data	± 0.013	± 0.014	± 0.013	± 0.013	± 0.013	± 0.013
	MC	± 0.014	± 0.014	± 0.014	± 0.014	± 0.014	± 0.014
	Data/MC ratio	± 0.004	± 0.005	± 0.005	± 0.003	± 0.001	± 0.001
Top/Bottom	Data	± 0.007	± 0.006	± 0.012	± 0.008	± 0.009	± 0.008
	MC	± 0.015	± 0.014	± 0.014	± 0.016	± 0.037	± 0.006
	Data/MC ratio	± 0.006	± 0.014	± 0.002	± 0.006	± 0.021	± 0.010
Global EM scale factor	Data	$+0.005$	$+0.005$	$+0.005$	$+0.000$	$+0.000$	$+0.000$
	MC	-0.013	-0.013	-0.014	-0.008	-0.008	-0.008
	Data/MC ratio	$-$	$-$	$-$	$-$	$-$	$-$
	Data	$+0.004$	$+0.004$	$+0.004$	$+0.000$	$+0.000$	$+0.000$
	Data/MC ratio	-0.010	-0.010	-0.010	-0.006	-0.006	-0.006
Total	Data	$+0.033$	$+0.047$	$+0.042$	$+0.062$	$+0.055$	$+0.050$
	MC	-0.035	-0.049	-0.044	-0.063	-0.056	-0.051
	Data/MC ratio	± 0.039	± 0.047	± 0.042	± 0.033	± 0.060	± 0.052
	Data	$+0.015$	$+0.023$	$+0.012$	$+0.042$	$+0.030$	$+0.023$
	Data/MC ratio	-0.017	-0.025	-0.015	-0.042	-0.031	-0.024

An Example Splash Event

- Approximately 140m down the beamline, the LHC beam was allowed to hit a collimator producing a large amount of particles
- The time shift of the PMT pulse can be seen as the particles propagate through the detector