The Compact Muon Solenoid Experiment

Conference Report

Mailing address: CMS CERN, CH-1211 GENEVA 23, Switzerland

01 March 2020 (v3, 30 March 2020)

Calibration and Performance of the CMS Electromagnetic Calorimeter in LHC Run-II

Chia-Ming Kuo for the CMS Collaboration

Abstract

Many physics analyses using the Compact Muon Solenoid (CMS) detector at the LHC require accurate, high resolution electron and photon energy measurements. Excellent energy resolution is crucial for studies of Higgs boson decays with electromagnetic particles in the final state, as well as searches for very high mass resonances decaying to energetic photons or electrons. The CMS electromagnetic calorimeter (ECAL) is a fundamental instrument for these analyses and its energy resolution is crucial for the Higgs boson mass measurement. Recently the energy response of the calorimeter has been precisely calibrated exploiting the full Run-II data, aiming at a legacy reprocessing of the data. A dedicated calibration of each detector channel has been performed with physics events exploiting electrons from W and Z boson decays, and photons from π^0/η decays. This talk presents the calibration strategies that have been implemented and the excellent performance achieved by the CMS ECAL with the ultimate calibration of Run-II data, in terms of energy scale stability and energy resolution.

Presented at *CHEF2019 Calorimetry for the High Energy Frontier 2019*

PREPARED FOR SUBMISSION TO JINST

CHEF2019 - CALORIMETRY FOR THE HIGH ENERGY FRONTIER 2019

25-29 NOVEMBER 2019

FUKUOKA, JAPAN

Calibration and Performance of the CMS Electromagnetic Calorimeter in LHC Run-II

Chia Ming Kuo^a on behalf of CMS collaboration

*^aDepartment of Physics and Center for High Energy and High Field Physics, National Central University
Chung-li, 32001, Taiwan*

E-mail: chia-ming.kuo@cern.ch

ABSTRACT: Many physics analyses using the Compact Muon Solenoid (CMS) detector at the LHC require accurate, high resolution electron and photon energy measurements. Excellent energy resolution is crucial for studies of Higgs boson decays with electromagnetic particles in the final state, as well as searches for very high mass resonances decaying to energetic photons or electrons. The CMS electromagnetic calorimeter (ECAL) is a fundamental instrument for these analyses and its energy resolution is crucial for the Higgs boson mass measurement. Recently the energy response of the calorimeter has been precisely calibrated exploiting the full Run-II data, aiming at a legacy reprocessing of the data. A dedicated calibration of each detector channel has been performed with physics events such as electrons from W and Z boson decays, and photons from π^0/η decays. This paper presents the calibration strategies that have been implemented and the excellent performance achieved by the CMS ECAL with the ultimate calibration of Run-II data, in terms of energy scale stability and energy resolution.

KEYWORDS: Calorimeters, Calorimeter methods

Contents

1	Introduction	1
2	Energy reconstruction for the electromagnetic objects	2
3	Crystal response monitoring	2
4	Intercalibration and energy scale	3
5	ES calibrations	4
6	Supercluster correction and performance of CMS ECAL	4
7	Summary	5

1 Introduction

The Compact Muon Solenoid (CMS) experiment [1] at the LHC has a broad physics programme, ranging from measurements of the Standard Model (SM) and the recently discovered Higgs boson, to studies of heavy-ion collisions, to searches for new physics beyond the SM. Many physics analyses involve electrons or photons in the final state, and the electromagnetic calorimeter (ECAL) plays a crucial role in their reconstruction and identification.

The CMS ECAL [2] is a compact, hermetic, fine-grain, and homogeneous calorimeter made of 75848 lead-tungstate (PbWO_4) scintillating crystals, divided among a barrel region (EB) with $|\eta| < 1.48$ and two endcaps (EE) that extend up to $|\eta| < 3.0$. The high density ($\rho = 8.28 \text{ g/cm}^3$), short radiation length ($X_0 = 0.89 \text{ cm}$) and small Molière radius ($R_M = 2.19 \text{ cm}$) of the PbWO_4 crystals allow the construction of a compact calorimeter to fit inside the 3.8T superconducting solenoid with fine granularity. These crystals also feature fast light emission (80% of light emitted in $25 \mu\text{s}$). The scintillation signal is detected by two avalanche photodiodes (APDs) per crystal in the EB and one vacuum phototriode (VPTs) in the EE, respectively. These characteristics help provide precise energy, good timing resolution, and fast and efficient readout for online selection.

In addition, a preshower detector (ES) is placed in front of EE at $1.65 < |\eta| < 2.6$. ES has two layers in each endcap: lead absorbers ($2X_0$ and $1X_0$, respectively) initiate electromagnetic showers from incoming electrons/photons while silicon strip sensors with 1.9mm strip pitch placed after each absorber measure the energies deposited and the transverse shower profiles. The main purpose of the ES is to enhance the photon identification capabilities.

For the energy resolution of the CMS ECAL, the uniformity and stability are required to be better than 0.5%. In EB, 1% energy resolution was achieved for unconverted and late-converting photons in the LHC Run-I [3]. During the Run-II (2015-2018) [4], the LHC operated at an increased center-of-mass energy of 13 TeV with a challenging environment: an average of 34 proton-proton

interactions per bunch crossing (pileup), reduced bunch spacing of 25 ns, and increased radiation level. To overcome these challenges, novel reconstruction algorithms and calibration procedures were developed to maintain the excellent performance of the CMS ECAL. The optimal performance was achieved by exploiting full statistics of each data-taking year.

2 Energy reconstruction for the electromagnetic objects

For unconverted photons, on average about 70% of their energy deposits in a single crystal and 97% in a 3×3 array. However, due to the presence of tracker material in front of ECAL causing the radiation of energy (secondary emission by bremsstrahlung or photon conversions), the electromagnetic particles would deposit their energy over several PbWO_4 crystals. The electromagnetic showers spread along the ϕ direction because of the strong magnetic field of CMS. Dynamic clustering algorithms [5, 6] are used to collect the energy deposits in ECAL.

The reconstructed energy of electrons and photons is estimated by:

$$E_{e,\gamma} = F_{e,\gamma} \times [G(\eta) \times \sum_i (C_i \times S_i(t) \times A_i \times H_{i\eta}) + E_{ES}], \quad (2.1)$$

where the sum runs over all clustered crystals, labelled by i . The signal amplitude in ADC measured in i -th crystal is denoted as A_i . $S_i(t)$ is a time dependent correction that accounts for time variations of the channel response due to changes in crystal transparency. C_i , the intercalibration coefficients, are applied to equalize the channel response due to the differences of light yield and photodetector response among crystals. $H_{i\eta}$ corrects for the η -scale to equalize η ring response with respect to the simulation. $G(\eta)$ is a scale factor for the ADC-to-GeV conversion and η -dependence to incorporate geometry effects. E_{ES} is the energy measured by the ES. $F_{e,\gamma}$ is the cluster correction for imperfect clustering and other geometry effects, and it is different for photons and electrons.

The level of out-of-time pileup increases when the bunch spacing is reduced. This leads to a higher probability of multiple scintillation processes from preceding and trailing bunches starting in the same crystal causing the imprecise in-time amplitude reconstruction. The "multifit" reconstruction method [7], a template fit with multiple components, was developed to mitigate this effect and is used in both of the online and offline reconstruction in Run-II. An example of fitted pulses for a simulated event is illustrated in Fig. 1a. The energy resolution is improved with respect to the Run-I method [8], in particular for the electrons and photons with low p_T due to the larger fraction of deposits from pileup in the total energy.

3 Crystal response monitoring

Radiation damage to the crystals causes decreases in the crystal response due to loss of transparency, and variations in the VPT response in the EE. However, the response can increase through thermal annealing in the absence of radiation. These effects lead to the need of time dependent corrections, $S_i(t)$, on short time-scales.

A dedicated monitoring system [9] was designed to inject laser light with a wavelength of 447 nm into each crystal. This system provides corrections for the variations in response described above. It measures the calibration point per crystal every 40 minutes and corrections are obtained

Figure 1: (a) An example of fitted pulses for a simulated event is shown for a signal in the EE. (b) History of the relative response to laser light injected in the PbWO₄ crystals measured by the laser monitoring system.

and applied in about 48 hours for the prompt reconstruction. The evolution of relative response, averaged over all crystals in bins of pseudorapidity η , from 2012 – 2018 is summarized in Fig. 1b. The response changes are up to about 10% in the EB and reach up to about 60% at $|\eta| = 2.5$, the limit of the tracker acceptance. The change is up to 96% in the regions closest to the beam pipe. Damage and recovery cycles are remarkable, as well as steady recovery periods during the LHC stops. The response has not fully recovered, however, especially in the high $|\eta|$ regions. The validation of response monitoring is performed using the stability of the ratio of the energy (E) measured by the ECAL to the momentum (p) measured by the tracker with the isolated electrons decaying from W or Z bosons. The stability of the π^0 invariant mass, m_{π^0} , as a function of time is also employed to monitor and certify the corrections.

4 Intercalibration and energy scale

The relative calibration between crystals, referred to as intercalibration, is performed to ensure uniformity across the detector. Several techniques are exploited to derive the refined intercalibration factors, C_i , *in situ* using collision events. The π^0 method constraints the invariant mass of unconverted photon pair of π^0 . The photons are reconstructed as the matrices of 3×3 crystals in this technique. The C_i of the central crystal of the matrix is updated in each iteration based on the fitted m_{π^0} and the iterative procedure stops when the fitted m_{π^0} converges to the PDG value of m_{π^0} . In the method of E/p ratio, the isolated electrons from W and Z decays are utilized and the iterative procedure is also used to minimize the spread of the E/p distribution. In the $Z \rightarrow ee$ method, a fit is performed to derive the C_i constants from the invariant mass of the electron pair.

The precision of each intercalibration method is evaluated with the relative energy resolution of $Z \rightarrow ee$. The precision from the various methods with 2018 data as a function of η is shown in Fig. 2a. The black points represent the precision of the combination of the three methods, obtained from the mean of the individual method constants at a fixed η bin, weighted by their respective precision. The combination yields a precision of around 0.3% for the central region of the EB, $|\eta| < 1.1$, and better than 1.2% for $|\eta|$ up to 3.

$H_{i\eta}$ is obtained from the peak of the invariant mass of $Z \rightarrow ee$ events, selecting a sample of electrons with low emission of bremsstrahlung, to equalize the relative response among different η regions. The η scale is set to match the expectation in the simulation. The overall energy scale $G(\eta)$ is determined separately for the EB and EE, such that the peak of reconstructed Z mass in both of data and simulation agrees.

5 ES calibrations

To calibrate ES, particles with momentum greater than 1 GeV are employed. The majority is $\pi^\pm$ and they are nearly minimum ionizing, so for simplicity they are referred to as "MIPs". The MIP calibration is derived with special data taken by setting the front-end electronics with a good S/N in order to see the MIP peak for every 10-20 fb^{-1} . A Landau convoluted with a Gaussian is fitted to the energy spectrum of each silicon channel to extract the calibration parameter for each silicon sensor. The design-goal accuracy of the channel-by-channel calibration is set to 5%. This corresponds to a contribution of about 0.25% to the overall EE+ES energy resolution for high energy electrons since only a few percent of a photon or electron energy is deposited in ES. The history of the MIP response is shown in Fig. 2b. The response decreases as a function of luminosity in general and faster in the higher η region. This implies the ES sensors are more affected by the radiation damage in the more forward region. The only exception is in the beginning of 2017 data taking, when the bias voltage was increased. Finally the measured energy of ES cluster is stabilized by applying a correction computed by minimizing the χ^2 value between the energy distribution of data and MC with $Z \rightarrow ee$ events.

Figure 2: (a) Precision of the intercalibration constants as a function of η derived from $\pi^0 \rightarrow \gamma\gamma$, E/p , $Z \rightarrow ee$ method, and their combination using 2018 data. (b) The evolution of ES MIP response with respect to the beginning of 2015 in the front layer.

6 Supercluster correction and performance of CMS ECAL

Finally the correction to the supercluster energy, $F_{e,\gamma}$, is obtained using a multivariate approach that maximally exploits the information of the events, such as shower shapes, shower position in ECAL local and CMS global coordinates, and global event variables sensitive to pileup effects. It is tuned based on the simulated events and then validated on data.

The response stability of the reconstructed Z mass and shower shape of electrons after corrections is constantly monitored using $Z \rightarrow ee$ events. The reconstructed Z mass is stable over time and its variation is found to be around 0.2%. The shower shape of electrons also shows stable response and the relative spread is better than 0.25%.

The relative energy resolution, σ_E/E , is extracted from $Z \rightarrow ee$ events using an unbinned likelihood fit to the invariant mass distribution with a Breit-Wigner function convoluted with a Gaussian as the signal model. The energy resolution with reprocessed data collected in 2018 after the refined calibration achieves a resolution that is better than 2.5% for $Z \rightarrow ee$ in the central EB and 4-5% in the EE.

In terms of physics analysis, the per-event relative mass resolution estimates are used to classify several "untagged" categories for the di-photon invariant mass fit in the analysis of $H \rightarrow \gamma\gamma$. A mass resolution of about $\sigma_m/m_{\gamma\gamma} = 1\%$ is obtained for the best category in Ref. [10]. This result is comparable to the Run-I performance despite more challenging conditions in Run-II.

7 Summary

The CMS ECAL had shown excellent performance throughout LHC Run-II. More challenging conditions during data taking in Run-II require optimized strategies for reconstruction and calibration. The stable response of energy and shower shape, good energy resolution, and excellent $H \rightarrow \gamma\gamma$ mass resolution comparable to Run-I were achieved.

References

- [1] S. Chatrchyan *et al.*, *The CMS experiment at the CERN LHC*, *JINST* **3** (2008) S08004.
- [2] The CMS Collaboration, *The Electromagnetic Calorimeter Project: Technical Design Report*, CERN/LHCC 1997, CMS TDR 4.
- [3] The CMS Collaboration, *Observation of the diphoton decay of the Higgs boson and measurement of its properties*, *Eur. Phys. J. C* **74** (2014) 3076.
- [4] J. Wenninger, *Operation and Configuration of the LHC in Run 2*, CERN-ACC-NOTE-2019-0007.
- [5] The CMS Collaboration, *Performance of electron reconstruction and selection with the CMS detector in proton-proton collisions at $\sqrt{s} = 8$ TeV*, *JINST* **10** (2015) P06005.
- [6] The CMS Collaboration, *Performance of photon reconstruction and identification with the CMS detector in proton-proton collisions at $\sqrt{s} = 8$ TeV*, *JINST* **10** (2015) P08010.
- [7] E. Di Marco, *CMS electromagnetic calorimeter calibration and timing performance during LHC Run I and future prospects*, CMS-CR-2014-410, <https://cds.cern.ch/record/1975982>.
- [8] P. Adzic *et al.*, *Reconstruction of the signal amplitude of the cms electromagnetic calorimeter*, *Eur. Phys. J. C* **46** (2006) 23.
- [9] M. Anfreville *et al.*, *Laser monitoring system for the CMS lead tungstate crystal calorimeter*, *Nucl. Instr. Meth. A*, **594** (2008) 292.
- [10] The CMS Collaboration, *A measurement of the Higgs boson mass in the diphoton decay channel*, arXiv:2002.06398 [hep-ex], *submitted to Phys. Lett. B*.