EUROPEAN ORGANIZATION FOR NUCLEAR RESEARCH

CERN - AB Division

CERN-AB-2003-018 ABP

Beam Photography : A technique for Imaging Dark Currents

P. Gruber, Y. Torun

Abstract

Dark currents are unwanted electron currents that stem from surface electrons that are emitted in rf cavities and accelerated in their electric field. This paper describes a novel technique to produce a 2D-image of dark currents and presents some results.

Ordinary black and white photographic paper has been exposed to dark current electrons stemming from an 805Mhz cavity in the Lab G facility at Fermilab. This cavity is a closed-cell cavity in a 2.5 T solenoidal field parallel to the electrical field. Thin rf and vacuum windows allow the electrons to exit the cavity. Due to the focusing effect of the magnetic field, a sharp picture of the sources of the dark currents is obtained. Single emitters, surrounded by regions of virtually no dark current, can be clearly identified.

*Presented at PAC 2003, Portland, Oregon, USA
from 12 to 16 May 2003*

*Geneva, Switzerland
22 May 2003*

BEAM PHOTOGRAPHY: A TECHNIQUE FOR IMAGING DARK CURRENTS

P. Gruber, CERN, Geneva*, Switzerland
Y. Torun, IIT, Chicago, IL, USA

Abstract

Dark currents are unwanted electron currents that stem from surface electrons that are emitted in rf cavities and accelerated in their electric field. This paper describes a novel technique to produce a 2D-image of dark currents and presents some results.

Ordinary black and white photographic paper has been exposed to dark current electrons stemming from an 805 MHz cavity in the Lab G facility at Fermilab. This cavity is a closed-cell cavity in a 2.5 T solenoidal field parallel to the electrical field. Thin rf and vacuum windows allow the electrons to exit the cavity. Due to the focusing effect of the magnetic field, a sharp picture of the sources of the dark currents is obtained. Single emitters, surrounded by regions of virtually no dark current, can be clearly identified.

INTRODUCTION

Dark currents are unwanted electron currents that stem from surface electrons that are emitted in rf cavities and accelerated in their electric field.

Dark currents have been studied in the context of the international Muon Ionization Cooling Experiment (MICE)[1], which is an accelerator R&D experiment for ionization cooling at a future neutrino factory[2]. Ionization cooling is a two step process in which muons are first sent through an absorber (mostly liquid hydrogen) in order to reduce both p_t and p_l . In the second step, the particles are re-accelerated in an electric field parallel to the beam axis and p_l is restored. The net effect is a reduction of $x' = \arctan p_t/p_l$ and – as the focusing is kept constant – a reduction of the normalized emittance.

MICE has three main components: an incoming spectrometer, the cooling apparatus with absorbers and cavities, and an outgoing spectrometer. The emittance before and after cooling is measured on a particle-by-particle basis in the spectrometers. The layout of MICE in Fig. 1 shows how close the spectrometers are to the cavities. In fact, they are only separated by the 35 cm of liquid hydrogen absorber. This makes the measurement vulnerable to dark currents and to x-rays produced by dark current electrons and explains the interest of MICE in controlling dark current emissions.

Figure 1: Layout of the MICE experiment. A differential measurement is performed between the two spectrometers. In the middle, there is the cooling apparatus with two cavities and three liquid hydrogen absorbers. The whole experiment is embedded in a fine-tuned set of solenoid magnets.

Frequency	805 MHz
Accelerating gradient	13 MV/m
Total accelerating voltage	1.17 MV
Magnetic field	2.5 T

Table 1: Operation parameters of the cavity and magnet.

EXPERIMENT SETUP

The MUCOOL collaboration operates a 805 MHz test cavity at the Fermilab Lab G[3]. This cavity has been specially designed to study dark currents and x-rays emitted from rf cavities. It features a pair of very thin windows, allowing the dark current electrons to exit the cavity. These windows can be changed. All measurements presented here have been performed with a 200 μm Cu rf window and a 200 μm Ti vacuum window.

The cavity is immersed in a superconducting solenoidal magnet with fields up to 5 T. All measurements except the one reported at the end were performed at a magnetic field of 2.5 T. The cavity operation parameters can be found in Tab. 1. To make the comparison easier, all measurements were performed at field gradient 13 MV/m.

METHODOLOGY AND MATERIAL

Ordinary photographic paper for black-and-white enlargements was used to detect the electrons. The paper was wrapped in contractor-grade black plastic foil to prevent exposure to ambient light. It was put perpendicular to the beam axis at different distances to the cavity window, exposed for 30 s to 9 hrs and subsequently developed. The exposure time was adapted to the expected flux.

The advantages of photographic paper are its moderate speed (slower than polaroid paper, yet fast enough to get results in minutes), the low cost, its high grain and the large sizes available. The details of the material are outlined in Tab. 2.

* peter.gruber@cern.ch

Figure 2: Distance series of photographic paper, taken at various distances of from the cavity window. The exposure time was 1 min for the first three pictures and 3 min for the latter two. All pictures are on the same scale.

After the development, the photographic paper was scanned using a Canon 20LIDE scanner with 300 dpi and 16 bits. To calibrate the scans, a sample white and black point was scanned with each picture. The levels of grey expressed at points of interest in percentage points were then extracted using a picture editing program. It was assumed that the darkening is roughly proportional to the dose rate for levels of grey of up to 75%.

DISTANCE MEASUREMENT SERIES

A series of pictures were taken on the solenoid axis at varying distances from the cavity windows. The results are shown in Fig. 2. The first picture was taken at a distance of 1 cm from the cavity window, well inside the solenoid. The second picture was taken at the solenoid's end. The other three pictures were taken in the fringe field of the solenoid at distances of 20 cm each.

The pictures show that the black spots seen on the first plane dilute and grow as the fringe field of the solenoid opens up. This picture makes it clear, that the signal seen is really induced by electrons and not by x-rays, which are also emitted from the cavity.

The last two pictures illustrate very well how the electrons follow the magnetic field lines of the fringe field and how they get diluted in this process. This is important for the MICE experiment, because there is a flip in the field direction between cavity and spectrometers for most configurations. As seen in Fig. 2, the dark current electrons follow the field lines and virtually none will reach the spectrometer.

EXPOSURE TIME SERIES

At a fixed position at the end of the solenoid (this compares to position 2 in Fig. 2) a time series of exposures was taken. The exposure times were: 30 s, 45 s, 1 min, 2 min and 5 min. The results, shown in Fig. 3, indicate a background rate that varies highly with the location. One can imagine dark current "beamlets" that stem each from a single emitter.

Combining different exposure times makes it possible to extend the dynamic range of the photographic paper. If one looks at the 30 s picture, there is spot that is completely black (marked with a dotted circle), while another area

Photographic paper	ILFORD® Multigrade IV, FB Fiber, MGF.5K [4] sizes 8x10 in and 11x14 in (20x25 cm and 27.5x35 cm)
Developer	Ilford Multigrade Paper developer, 1+9 diluted, 2 min
Fixer	Ilford Rapid fixer, 1+9 diluted, 2min

Table 2: Specifications of the photographic material used

(marked with a dotted triangle) is still completely white. After 10 times the exposure time (300s), the region marked with a triangle is 10% grey. If one spot is only 10% grey after 10 times the exposure time, one can conclude that different beamlets vary at least $10 \cdot 10 = 100$ times in intensity.

PHOTOGRAPHIC CALORIMETER

A photographic calorimeter was built with six layers of photographic paper separated by layers of 1.6 mm aluminum each. To be able to see high-energy particles with low rates and attenuated x-rays, the exposure time has been expanded to 9 hrs. It was thus expected that the first sheet of paper, which is not protected by the aluminum, would be completely black.

The picture series is shown in Fig. 4. The light grey ring is due to a mounting device, that cast a shadow on the calorimeter. The active area of sheet no. 0 is completely black, as expected. A certain number of electrons penetrates the first layer of Al and can be identified on sheet no. 1. No electrons are seen after the second layer of Al on sheets no. 2 and so forth; only a uniform x-ray background can be identified.

This result is perfectly compatible with a MARS[5] simulation, that shows that for 1.2 MeV electrons, 20% survive one Al layer, and less than 10^{-4} survive two layers. One can conclude that at least a large fraction of the observed electrons have been accelerated through the full length of the cavity.

The x-ray background shows the expected attenuation pattern, with slightly lower intensities after each layer of Al.

MEASUREMENTS WITHOUT MAGNETIC FIELD

For comparison purposes, one exposure was carried out with no magnetic field. The photographic paper was located at the end of the solenoid. The exposure time was extended to 30 min. No structure whatsoever could be detected. The exposure was 60%. This translates into an exposure time of 50 min for full black, a factor of 100 more than the darkest spots *with* magnetic field at the same location.

Figure 3: Time series of photographic paper, located at the end of the solenoid, exposed 30 s, 45 s, 1 min, 2 min and 5 min.

Figure 4: Image series of the photographic calorimeter. The first image was directly exposed, then there was a shielding of 1.6 mm Al between each plate. The exposure time was 9 hrs.

CONCLUSIONS

For the **beam photography method**, one can conclude:

- Beam photography is capable of producing accurate 2D-images of dark currents.
- This method works only, if a magnetic field parallel to the beam axis guides the dark current filaments.
- X-rays are normally not detected, as they produce a signal that is a factor of 500 weaker than the electrons' signal.

Following facts about **dark currents** produced by the Lab G cavity have been revealed:

- About 50 individual emitters have been identified.
- The intensity of individual dark current filaments varies by a factor of more than 100.
- A large fraction of the dark current particles has been accelerated along the whole length of the cavity.

Leading to these conclusions for the **MICE experiment**:

- A field flip at the location of the absorber solves the problem of low-energy electrons, as these follow the field lines and are lost.
- Without field flip, the dark current filaments follow the field lines and reach the spectrometer. There, the local background rate varies by a factor of more than 100. When analysing background effects in the spectrometers, it is not enough to consider the average background rate.

ACKNOWLEDGEMENTS

The authors wish to thank the MUCOOL collaboration as well as A. Moretti and J. Norem for their help. This work has in part been supported by the University of Geneva (CH) and the IIT (USA).

REFERENCES

- [1] D. Kaplan, this conference.
- [2] P. Gruber (ed.), The Study of a European Neutrino Factory Complex, CERN/PS/2002-080.
- [3] J. Norem, this conference.
- [4] http://www.ilford.com/html/us_english/prod_html/multifb/multi_iv_fb.html
- [5] N. V. Mokhov, The Mars Code System's User Guide, Fermilab-FN-628 (1995).